NEEDLE HOLDER

Instruction for use AA74

The intended purpose of Needle Holder is to hold and guide suture needles securely for suturing

TESTING AND INSPECTION:

Inspect use for rust, pits, crakes, pores & dents before use.

WARNINGS:

Don't use for second time

Don't use if you observe sharp edges

Product material is conductive take away from electric source.

Don't use without Sterilization don't use

Don't use wrong of instrument can be creating problem in operation

STORAGE & TRANSPORT CONDITIONS:

Ambient temperature range: 32°F (0°C) to 122°F (50°C) Relative humidity range: 30% to 75%

Atmospheric pressure range: 700 hPA to 1060 hPA

RECYCLE:

When reprocessing medical devices please follow your local government Health & Safety procedures

SURGIKARE
Toor Abad Daska Road,
Sialkot, Punjab - PK

EC REP
Obelis
Bd Général Wahis, 53
B-1030 Brussels - BE

CE 1639

Doc. # IFU-12-05/00

Internal drafting date 16/10/2026